

## Cat No. AID 025

### Intended use

CP4EPSPS (RR) lateral flow strips (LFS) test kit is intended to be used for the qualitative detection of CP4EPSPS protein in individual Corn/Soybean leaf or seed samples.

The total incubation time of the assay is 5 minutes.

### Principle of the test

An antibody specific to the CP4EPSPS protein molecule is immobilized on the test line area of nitrocellulose membrane. Second antibody specific for CP4EPSPS molecule is conjugated with colloidal gold and incorporated in the sample area of lateral flow strip. Anti mouse IgG is immobilized on the control line area of the LFS.

When the LFS is placed in the sample extract, the CP4EPSPS protein present in the sample extracts binds to the antibody labelled with gold nano particles and the complex moves upward by capillary action. The complex then binds to the antibody coated on the test line resulting in pink/purple color test line. As the complex moves further up, it binds to the control line resulting in pink/purple color control line. In absence of CP4EPSPS, the test line does not appear as no complex binds to the test line while control line turns pink/purple color indicating validity of test protocol.

### Cross-reactivity

The CP4EPSPS LFS kit does not cross react with Cry1F, Cry1Ac, eCry3.1Ab (AgriSure Duracade), Cry3Bb, Cry1Ab, Cry2A, Vip3A, Cry34Ab1, Cry35Ab1 or PAT (LibertyLink\*\*\*)

### Contents of the kit: Kit is sufficient for 100 tests

CP4EPSPS LFS strips	50 strips per canister. Two canisters per kit
Extraction buffer	Two vials of 50 ml each. Ready to use
Dropper	One
Pack insert	One

### Material and equipment required but not provided

- ❖ Pair of scissors
- ❖ Pestles
- ❖ Marking pen and Paper towels.
- ❖ 1.5 ml Microfuge tube with cap
- ❖ Timer

### Precautions

The CP4EPSPS LFS kit is intended for in vitro use only. The reagents contain sodium azide as preservative. Prevent direct skin and eye contact with kit components. Seek medical attention in case of accidental ingestion of kit components.

### Storage of the kit

The kit should be stored at 2 - 8° C. The unopened kit is stable till the expiry date printed on the kit label. The cap of the canister should be closed firmly after removing the required strips. Exposure to moisture is likely to affect the performance of the test strips.

The kit should not be frozen.

### Sample preparation

#### **Extraction of seed tissue :**

Crush single seed and transfer it to 1.5ml microfuge vial . Add 1.0 ml of extraction buffer. Mix well and wait for five minutes.

#### **Extraction of leaf tissue:**


Take two leaf punch (approximate weight 20 mg) and transfer it to 1.5 ml microfuge tube. Grind the leaf tissue with pestle until the leaf tissue is well ground. Add 0.5 ml extraction buffer. Mix well and wait for 5 minutes. Use new pestle for each leaf sample to avoid cross contamination.




Extracting protein by crushing with pestle

## Assay Procedure

1. Allow canister to come to room temperature before opening it to remove the desired number of strips.
2. Insert one strip in each sample. Part of the strip showing arrow should be dipped in sample extract. Allow the test strip to remain in the seed crushing plate/microfuge tube in vertical position for 5 minutes.
3. Remove the strips and observe the results. Positive sample results may appear much earlier than 5 minutes.
4. For permanent storage of strips, cut off the bottom section of the strip covered with arrow using pair of scissors.



LFS Strip in Sample



Cutting of strip for permanent record and storage

## Interpretation of LFS results

Read the strip in 5 minutes.

Two line indicates positive test result while single line indicates negative results. Absence of control line in 5 minutes indicates invalid test.

The appearance of faint test line after 5 minutes should not be necessarily interpreted as positive test.

## Notes

- ❖ The procedure instructions should be strictly followed to get correct results. Change in procedure may lead to wrong results.
- ❖ This kit has been optimized for individual Corn/Soybean seed and leaf.
- ❖ Cross contamination between different microfuge tubes will lead to wrong results
- ❖ This kit is meant for screening for presence or absence of CP4EPSPS in individual Corn/ Soybean seed and leaf.
- ❖ When in doubt, please confirm results with an alternate method.

## WARRANTY

Eurofins Amar Immunodiagnostics Pvt. Ltd. warrants that the products sold hereunder (“the Product”) are defect-free in material and workmanship, provided they are used in accordance with the prescribed instructions before the expiry of the products as printed on the product label.

The customer should notify Eurofins Amar Immunodiagnostics in writing of Warranty defects during the warranty period, including an offer by the customer to return the Products to Eurofins Amar Immunodiagnostics for evaluation. Eurofins Amar Immunodiagnostics will repair or replace, at its sole option, any product or part thereof that proves defective in materials or workmanship within the warranty period.

This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Eurofins Amar Immunodiagnostics

## THIS WARRANTY IS EXCLUSIVE

The sole and exclusive obligation of Eurofins Amar Immunodiagnostics shall be to repair or replace the defective Products in the manner and for the period provided above.

Eurofins Amar Immunodiagnostics shall not have any other obligation or liability, whatsoever it may be, with respect to the Products or any part thereof. Under no circumstances, whatsoever the circumstances may be, shall Eurofins Amar Immunodiagnostics be liable for incidental, special, or consequential damages.

If any part of this Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

\*Parafilm is a registered trademark of American Can Corporation (now Pechinney Plastic Packaging).

\*\*Bollgard & Roundup Ready are registered trademarks of the Monsanto Company.

\*\*\*Liberty Link is a register trademark of Bayer Crop Science.

....

**For any further information, please contact:**

**Eurofins Amar Immunodiagnostics Pvt. Ltd.**

242/1, Road No. 18, Jubilee Hills, Hyderabad - 500033 India

Tel : +91 40 2355 2953

E-mail : [technologies.amar@eurofins.com](mailto:technologies.amar@eurofins.com)

Website : [www.eurofins-technologies.com](http://www.eurofins-technologies.com)