

## INSTRUCTION FOR USE

### ELISA for detection of Cry1Ac and Cry2A

Cat No. AID 017



## 1 INTENDED USE

The Bollgard \*\*II combination kit is used for simultaneous qualitative detection of Cry1Ac and Cry2A protein in cotton leaf or seed samples with both analytes, measured in the same well of the assay plate. The format helps the laboratory to cut down work load by half thereby increasing productivity. The kit is ideally suited for testing large numbers of seed lots for Cry1Ac and Cry2A simultaneously.



## 2 PRINCIPLE OF THE TEST

In the test, cotton sample extracts are added to test wells coated with antibodies raised against Cry1Ac and Cry2A proteins. Any Cry1Ac or Cry2A protein present in the sample extract binds to the antibodies. This is detected by addition of alkaline phosphatase - labeled Cry2A antibody or horseradish peroxidase – labeled Cry1Ac antibody. After a simple wash step, the results of the Cry2A assay are visualized via the addition of a pNPP substrate. Once the yellow color develops and is read, the wash step is repeated, and TMB substrate is added. The Cry1Ac results are visualized via the development of the resulting blue color which, subsequently, becomes yellow on addition of stop solution.

## 3 CROSS-REACTIVITY

The Cry1Ac - Cry2A ELISA Kit does not recognize CP4EPSPS (Roundup Ready\*\*), Cry1F, Cry9C, CryBb, mCry3A, Cry1C or PAT (Liberty Link\*\*\*).

## 4 CONTENTS OF THE KIT:

### **Kits are provided in 1 plate format.**

- One plate of 96 wells coated with Anti-Cry1Ac & Cry2A antibody, each packed individually in a laminate bag with silica gel
- Cry1Ac & Cry2A combined conjugate: One ready – to - use bottle of 6 ml
- Extraction buffer concentrate: (10X): 50 ml, 1 vial.
- Wash solution (10X): 50 ml, 1 vial.
- Substrate #1 (pNPP): One ready-to-use bottle of 12 ml
- Substrate #2 (TMB): One ready-to-use bottle of 12 ml

### **Kits are provided in 10 plates format.**

- Ten plates of 96 wells coated with Anti-Cry1Ac & Cry2A antibody, each packed individually in a laminate bag with silica gel
- Cry1Ac & Cry2A combined conjugate: One ready – to - use bottle of 55 ml
- Extraction buffer concentrate: One packet of powder with a 25ml vial of Tween-20
- Wash solution concentrate: One packet of powder with a 2.5 ml vial of Tween-20
- Substrate #1 (pNPP): One ready-to-use bottle of 110 ml

### **Kits are provided in 50 plates format.**

- Fifty plates of 96 wells coated with Anti-Cry1Ac & Cry2A antibody, each packed individually in a laminate bag with silica gel
- Cry1Ac & Cry2A combined conjugate: One ready – to - use bottle of 275 ml
- Extraction buffer concentrate: One packet of powder with a 25ml bottle of Tween-20
- Wash solution concentrate: One packet of powder with a 2.5 ml bottle of Tween-20
- Substrate #1 (pNPP): One ready-to-use bottle of 500 ml

## ELISA for detection of Cry1Ac and Cry2A

Cat No. AID 017



- |  |   |   |
|--|---|---|
| ml   | ml  | ml  |
| • Cry1Ac / Cry2A Positive control: One ready-to-use vial of 2 ml | • Substrate #2 (TMB): One ready-to-use bottle of 110 ml           | • Substrate #2 (TMB): One ready-to-use bottle of 500 ml           |
| • Cry1Ac & Cry2A Negative control: One ready-to-use vial of 2 ml | • Cry1Ac / Cry2A Positive control: One ready-to-use vial of 10 ml | • Cry1Ac / Cry2A Positive control: One ready-to-use vial of 10 ml |
|  | • Cry1Ac & Cry2A Negative control: One ready-to-use vial of 10 ml | • Cry1Ac & Cry2A Negative control: One ready-to-use vial of 10 ml |

## 5 MATERIAL AND EQUIPMENT REQUIRED BUT NOT PROVIDED

- Pipette with disposable plastic tips
- Multichannel pipette with disposable pipette tips
- Plate ELISA reader with 450/620 nm filter
- Table top centrifuge
- Marking pen, Parafilm, Pipette tips, Timer and Paper towels
- Deionized or distilled water
- Graduated cylinders with one liter capacity
- Reagent troughs
- Plate shaker
- Plate ELISA washer or wash bottle

## 6 PRECAUTIONS

The Cry1Ac-Cry2A ELISA Kit is intended for *in-vitro* use only. The reagents contain Thimerosal as a preservative. Prevent direct skin and eye contact with kit components. Seek medical attention in case of accidental ingestion of kit components.

## 7 STORAGE OF THE KIT

1. The kit should be stored at 2 - 8° C. The unopened kit is stable till the expiry date printed on the kit label.
2. The expiry date of each unopened component is printed on the label of the component.
3. Bring all the components of the Kit to room temperature prior to use.

## 8 REAGENT PREPARATION

### 1. Preparation of Working extraction buffer:

- Dissolve the packet of extraction buffer concentrate powder in 500 ml of distilled/deionised water and add 25 ml of Tween-20 which is provided along with the powder. Mix it well before use. The resultant extraction solution is 10X concentrate.
- This has to be further diluted with 4500 ml of water to have a working extraction buffer.
- Store unused portion of working extraction buffer at 2 - 8° C. Thaw to room temperature before using the same again.

### 2. Preparation of Working wash buffer:

- Dissolve the packet of wash buffer concentrate powder in 500 ml of distilled/deionised water and add 2.5 ml of Tween-20 which is provided along with the powder. Mix it well

## ELISA for detection of Cry1Ac and Cry2A

Cat No. AID 017



before use. The resultant wash solution is 10X concentrate.

- This needs to be further diluted with 4500 ml of water to have working wash buffer
- Store unused portion of working wash buffer at 2 - 8° C. Thaw to room temperature before using the same again.

### 3. Preparation of stop solution:

- Add 27.0 ml of 98 % H<sub>2</sub>SO<sub>4</sub> in 973 ml of deionized or distilled water. Work in fume hood while preparing stop solution. Store at room temperature for up to one year.

### 4. Preparation of user positive control:

- The user is advised to prepare in-house positive control. Crush or grind one known positive seed into a uniform powder and add 1.0 ml of extraction buffer. Wait for 10 minutes before use. Prepare fresh before use.

## 9 SAMPLE PREPARATION

### 1. Individual seed

- Crush or grind the seed thoroughly into a uniform powder using mortar and pestle or a seed crusher. Add 1.0 ml of extraction buffer to each sample. Mix well and allow the mixture to stand for 30-60 minutes at room temperature. Allow particles to settle and use only the supernatant to do the test. Wash and rinse grinding equipment carefully between samples to avoid cross contamination.

### 2. Individual leaf

- Weigh 20 mg (about two leaf punch) and put it in a microtube. Crush leaf with pestle in 0.5 ml working extraction buffer. Mix well and incubate for 30-60 minutes at room temperature. Allow particles to settle and use only the supernatant to do the test.

Avoid cross contamination between samples.

## 10 ASSAY PROCEDURE

Test protocol for individual seed or leaf samples. Estimated procedure time is 60 minutes.

Test protocol for individual seed or leaf samples. Estimated procedure time is two hours (60+30+30 minutes)

1. Add 50 µl Cry1Ac-Cry2A enzyme conjugate in each well.
2. Add 50 µl of working extraction buffer in blank well.  
Add 50 µl of Cry1Ac-Cry2A positive control in two separate wells. Add 50 µl of Cry1Ac-Cry2A negative control in two separate wells. Add 50 µl of each sample to rest of the wells.
3. Incubate for 60 minutes at room temperature.
4. Remove content of the wells by decanting into a sink or a waste container. Add 300 µl/well wash solution to all wells and then empty wells by inverting the plate. Repeat washing procedure three more times.  
Alternately, perform four washes by using microtiter plate washer. After last wash, tap the inverted microtiter plate on paper towel to remove as much liquid as possible.
5. Add 100µl Substrate # 1.
6. Incubate for 30 minutes at room temperature.
7. Take absorbance at 405 nm with 630 nm as secondary filter. Presence of yellow color indicates that samples are positive for Cry2A.
8. Wash the plate in the same manner as indicated in step 4 above.
9. Add 100 µl Substrate # 2.
10. Incubate for 30 minutes at room temperature. Add 100 µl stop solution. Take absorbance at 450 nm with 630 nm as secondary filter. Presence of yellow color indicates that sample are positive for Cry1Ac.

## ELISA for detection of Cry1Ac and Cry2A

Cat No. AID 017



### Notes on Technique:

1. Protect the plates from draught, strong light or direct sunlight during the test procedure.
2. Careful aspiration of the washing solution is essential for good assay precision.
3. Since the timing of the incubation steps is important to performance of the assay, use multi channel pipettes to dispense samples/reagents.
4. Plate readers measure absorbance vertically. Do not touch the bottom of the wells.

## 11 INTERPRETATION OF RESULTS

### General test criteria:

1. Absorbance of the blank well should not exceed 0.1.
2. Absorbance value of blank should be subtracted from positive control, negative control and samples.
3. The coefficient of variance between the duplicate positive control should be less than 15 %
4. Mean positive control should have absorbance of at least 0.5.
5. Mean negative control should have absorbance below 0.1.
6. If above criteria are not met, the test is invalid and should be repeated.

### Qualitative Test

1. For single corn/cotton leaf and seed samples:  
**CUTOFF CALCULATION:** for single corn/cotton leaf or seed sample Cutoff value: (Mean absorbance of negative control + 0.1)
2. **Positive sample:** If sample absorbance is more than the cutoff value.
3. **Negative sample:** If sample absorbance is less than the cutoff value.
4. Low level results normally indicate improper washing, cross contamination, or technique. In such cases, retesting is recommended.

Performance characteristics: Sensitivity: 100 %; Specificity: 100 %

## 12 WARRANTY

Gold Standard Diagnostics Hyderabad warrants that the products sold hereunder (“the Product”) are defect- free in material and workmanship, provided they are used in accordance with the prescribed instructions before the expiry of the products as printed on the product label. The customer should notify Gold Standard Diagnostics Hyderabad in writing of Warranty defects during the warranty period, including an offer by the customer to return the Products to Gold Standard Diagnostics Hyderabad for evaluation. Gold Standard Diagnostics Hyderabad will repair or replace, at its sole option, any product or part thereof that proves defective in materials or workmanship within the warranty period.

This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Gold Standard Diagnostics Hyderabad.

## 13 THIS WARRANTY IS EXCLUSIVE

The sole and exclusive obligation of Gold Standard Diagnostics Hyderabad shall be to repair or replace the defective Products in the manner and for the period provided above. Gold Standard Diagnostics Hyderabad shall not have any other obligation or liability, whatsoever it may be, with respect to the Products or any part thereof. Under no circumstances, whatsoever the circumstances may be, shall Gold Standard Diagnostics Hyderabad be liable for incidental, special, or consequential damages. If any part of this Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

## ELISA for detection of Cry1Ac and Cry2A

Cat No. AID 017



\*Parafilm is a registered trademark of American Can Corporation (now Pechinney Plastic Packaging).

\*\*Bollgard & Roundup Ready are registered trademarks of the Monsanto Company.

\*\*\*Liberty Link is a register trademark of Bayer Crop Science.

### TECHNICAL SUPPORT SERVICE

For technical assistance and more information please contact the Eurofins Amar Immunodiagnostics Pvt. Ltd.

Customer Service or your local distributor.

#1504, 15<sup>th</sup> Floor, Manjeera Trinity Corporate, eSeva Lane,  
KPHB Phase 3, Kukatpally, Hyderabad, Telangana – 500072, India.

Tel: +91 7337480355

E-mail: [technologies.amar@eurofins.com](mailto:technologies.amar@eurofins.com)

[www.goldstandarddiagnostics.com](http://www.goldstandarddiagnostics.com).