



Effective Date: 23-FEB-2024

Hektoen Enteric Agar - Instructions for Use

Intended Use

BAC*Gro*TM Hektoen Enteric Agar (HE) is a selective and differential medium used in the isolation and differentiation of Gram-negative enteric pathogens.

Product Summary

Hektoen Enteric agar utilizes bile salts to inhibit Gram-positive organisms and some Gram-negative strains. Optimized differentiation is achieved through the addition of three carbohydrates: lactose, sucrose, and salicin. Ferric ammonium citrate and sodium thiosulfate allow for the detection of hydrogen sulfate production, aiding in differentiation by causing black-centered colonies for Salmonella. Acid fuchsin and Bromothymol blue is an indicator system with a lower toxicity to promote improved enteric pathogen recovery.

Formulation* (per Liter)

| Animal Peptone | 12.0 g |
|-------------------------|----------|
| Yeast Extract | 3.0 g |
| Bile Salts | 9.0 g |
| Lactose | 12.0 g |
| Sucrose | 12.0 g |
| Salicin | 2.0 g |
| Sodium Chloride | 5.0 g |
| Sodium Thiosulfate | 5.0 g |
| Ferric Ammonium Citrate | 1.5 g |
| Bromothymol Blue | 0.064 g |
| Acid Fuchsin | 0.1 g |
| Agar | 13.5 g |
| Total | 75.2 g/L |

^{*}Formula may be supplemented and/or adjusted as required to meet performance criteria

Directions

- 1. Add 75g of HE powder to 1 L of deionized water.
- 2. Stir while heating. Bring to a brief boil to dissolve completely.
- 3. DO NOT AUTOCLAVE.
- 4. Pour plates and allow to solidify.

Precautions

This product is for laboratory use only and should only be used by qualified, trained laboratory personnel. Personnel should always use proper aseptic technique and observe all biohazardous precautions. All microbiological cultures should be presumed to be infectious.

Avoid ingestion, inhalation, or contact with skin and mucous membranes. If contact occurs, flush the area with clean water.

Quality Control Specifications

Gold Standard Diagnostics tests each lot of manufactured BAC*Gro*TM culture media utilizing appropriate control organisms and specifications as documented on the Certificate of Analysis. End users should perform quality control testing in accordance with government regulatory requirements and accreditation guidelines. The following specifications are routinely used for testing:

Appearance (dehydrated): Green-beige, free-flowing, homogenous. May contain dark particles.

Appearance (prepared): Green-blue, slightly opalescent.

pH (prepared): 7.4 - 7.8 at 25°C

Organism Performance:

| Strain ID | Inoculum | Incubation | | | Result |
|--------------------------------------|----------|-------------|-------|-------------|---|
| | | Time | Temp. | Environment | |
| Salmonella Typhimurium (ATCC® 14028) | ≤100 CFU | 18 - 24 hr. | 35° C | Aerobic | Blue – blue- green CFU w. Black Centers |
| Shigella flexneri (ATCC® 12022) | ≤100 CFU | 18 - 24 hr. | 35° C | Aerobic | Green – green- blue CFU. |

| Escherichia coli | | | | | Partial/complete |
|---------------------------|----------------------|-------------|-------|---------|------------------|
| /ATCO® 25222\ | >10 ⁴ CFU | 18 - 24 hr. | 35° C | Aerobic | inhibition. |
| (ATCC® 25922) | | | | | Salmon CFU. |
| E. faecalis (ATCC® 29212) | >10 ⁴ CFU | 18 - 24 hr. | 35° C | Aerobic | Partial/complete |
| | | | | | inhibition. |
| | | | | | Yellow CFU. |

Limitations of the Procedure

This product is not labeled for use as a medical device, and is not intended to diagnose, treat, or prevent disease.

Due to variation in nutritional requirements, some species or strains may be encountered that grow poorly in this medium.

Further biochemical or serological testing is required for the identification of organisms grown in this medium.

Storage and Expiration

BAC Gro^{TM} HE should be stored at 2 – 30 degrees Celsius. Because of the hygroscopic nature of dehydrated culture media, it should be stored in a dry place and the lid should remain tightly sealed. Media should be discarded if it is not free flowing or shows discoloration.

The expiration date printed on the label is applicable to media stored as directed.

Catalog Numbers

DCM4901 – Hektoen Enteric Agar, 500g

DCM4910 - Hektoen Enteric Agar, 10kg

Effective Date: 23-FEB-2024

Revision History:

| Revision | Description | Effective Date |
|----------|---|----------------|
| 02 | Increased the inhibitory inoculum from 1,000 CFU to 10 ⁴ CFU to align with industry standards. Added in packaging part number DCM4910 | 23-FEB-2024 |
| 01 | Document creation | 13-APR-2021 |