


# Microcystin-DM ELISA Kit, Detailed Procedure


## 1. Addition of Standards, Samples

Add 100  $\mu$ L of the standard solutions, control or samples into the wells of the test strips according to the working scheme given. We recommend using duplicates or triplicates.




## 2. Addition of Enzyme Conjugate

Add 50  $\mu$ L of the enzyme conjugate to the individual wells successively using a multi-channel pipette or a stepping pipette.




## 3. Addition of Antibody Solution

Add 50  $\mu$ L of the Microcystin Monoclonal antibody solution to the individual wells successively using a multi-channel pipette. Cover the wells with parafilm or tape and mix the contents by moving the strip holder in a rapid circular motion on the benchtop. Be careful not to spill contents. Incubate the strips for 90 min. at room temperature.




## 4. Washing of Plates

After incubation, remove the covering and vigorously shake the contents of the wells into a sink. Wash the strips three times with a multi-channel pipette or using the diluted 1X washing buffer solution. Please use at least a volume of 250  $\mu$ L of washing buffer for each well and each washing step. Remaining buffer in the wells should be removed by patting the plate dry on a stack of paper towels.




## 5. Addition of Substrate/Color Solution

Add 150  $\mu$ L of substrate/color solution to the individual wells successively using a multi-channel pipette or a stepping pipette. Cover the wells with parafilm or tape and mix the contents by moving the strip holder in a rapid circular motion on the benchtop. Be careful not to spill contents. Incubate the strips for 20 min. at room temperature.




## 6. Addition of Stopping Solution

Add 100  $\mu$ L of stop solution to the wells in the same sequence as for the substrate solution using a multi-channel pipette or a stepping pipette.



## 7. Measurement of Color

Read the absorbance at 450 nm using a microplate ELISA reader. Calculate results.




For Ordering or Technical Assistance Contact:  
Eurofins Abraxis  
124 Railroad Drive, Warminster, PA 18974 USA  
Phone: 215-357-3911 Fax: 215-357-5232  
E-mail: [info.ET.Warminster@eurofinsUS.com](mailto:info.ET.Warminster@eurofinsUS.com)  
[www.abraxiskits.com](http://www.abraxiskits.com)

# Microcystin-DM ELISA Kit, Concise Procedure


## 1. Addition of Standards, Samples

Add 100  $\mu$ L of standard solutions, control or samples.




## 2. Addition of Enzyme Conjugate

Add 50  $\mu$ L of enzyme conjugate.




## 3. Addition of Antibody Solution

Add 50  $\mu$ L of the antibody solution. Cover and mix for 30 seconds by rotating on benchtop. Incubate for 90 minutes at room temperature.




## 4. Washing of Plates

Wash the plates three times with 250  $\mu$ L of diluted 1X washing buffer.




## 5. Addition of Substrate/Color Solution

Add 150  $\mu$ L of substrate/color solution. Incubate 20 minutes at room temperature and away from direct sunlight.




## 6. Addition of Stopping Solution

Add 100  $\mu$ L of stop solution.



## 7. Measurement of Color

Measure color at 450 nm. Calculate results.



For Ordering or Technical Assistance Contact:  
Eurofins Abraxis  
124 Railroad Drive, Warminster, PA 18974 USA  
Phone: 215-357-3911 Fax: 215-357-5232  
E-mail: [info.ET.Warminster@eurofinsUS.com](mailto:info.ET.Warminster@eurofinsUS.com)  
[www.abraxiskits.com](http://www.abraxiskits.com)